MANAMED®

ManaFlexx 2
NMES Lead Wires

48" Long Bifurcated Lead Wires
3mm Male Snap Connector
2mm Pin Connection
For Electrical
Stimulation

Part No. MF002LW
48" Standard, 1 Paired Wire

2022-06

Manufactured for:
ManaMed, LLC
2612 Sirius Dr.
Denton, TX 76208
Ph: 888-508-0712
www.ManaMed.com
Made in China

UDI

(01)00810113683225
(10)123456

1. Remove lead wires from package.
2. Connect snap end to the back of the ManaFlexx 2 unit.
3. Insert contact pins from lead wires into electrode receptacles.
4. Start treatment: Follow medical professional instructions and the Manaflexx 2 Instructions For Use.
5. After treatment remove electrodes from treatment site. Do not disassemble lead wires. Follow care instruction of electrodes.

Warnings and Precautions

⚠ For all relevant hazards, contraindications, warnings, precautions and adverse side effects, refer to the ManaFlexx 2 Instructions For Use.

- Keep electrode wires out of reach of children.
- Electrodes are intended for single person use only.
- DO NOT remove electrode by pulling on lead wire.
- Use only electrodes that fit the pin connector.
- Use only electrodes approved for NMES/TENS electrical stimulation.

Please read the Instructions For Use for the ManaFlexx 2 thoroughly before use.

ManaFlexx 2
NMES Lead Wires

Part No: **MF002LW**

Contents:
NMES Lead Wires

Suggested
HCPC: A4595

LOT 123456

REV A

Part No: **MF002LW**

Single Patient Multiple Use

LATEX

℞ only

MD

Consult instructions for use

NOT STERILE

KEEP DRY

UDI

(01)00810113683225
(10)123456

Part No: **MF002LW**

Sugg. HCPC: A4595
ManaFlexx 2 NMES Lead Wires