MANAMED

ManaFlexx 2
Electrode Sheet

Flexible, Reusable Self-Adhering

Electrodes For Electrical Stimulation

Part No. MFELP10

2"x 2" Standard, 10 Pack

2022-06

Manufactured for:
ManaMed, LLC
2612 Sirius Dr.
Denton, TX 76208
Ph: 888-508-0712
www.ManaMed.com
Made in China

UDI

(01)008101136PROOF
(10)123456

1. Clean skin thoroughly of any oils, lotions, dirt, etc. Apply electrode only to normal, intact, clean, healthy skin.
2. Remove electrodes from package and place at the treatment site as instructed. Press firmly on electrode to ensure a good contact.
3. Insert the contact pins from the device leads into the electrode receptacles.
4. Electrodes should be removed immediately after treatment.
5. Remove the electrode by pulling gently from the outer edge of the adhesive. Do not pull from the wire.
6. Return electrodes on the indicated side of the clear liner. Store the electrodes in the original packaging, keep and use at normal room temperatures. (68º–82ºF)
7. When adhesion diminishes, apply a few drops of water to the gel area and let absorb for one minute. Replace electrodes when this method becomes ineffective or they fail to stick firmly to the skin.
8. Electrodes should be used by one patient only.
9. Do not clean electrodes with anything other than plain tap water.

Warnings and Precautions

⚠ For all relevant hazards, contraindications, warnings, precautions and adverse side effects, refer to the ManaFlexx 2 Instructions For Use.

- Keep electrodes out of the reach of children.
- Electrodes are intended for single person use only.
- DO NOT remove electrode by pulling on lead wire.
- DO NOT use electrodes that are the wrong shape or size for ManaFlexx 2, or incorrectly applied. This could result in skin irritation or thermal burns.
- Keep electrodes separated during treatment.
- DO NOT apply to broken skin. Should a skin rash or skin burn occur, immediately discontinue use and contact your physician.
- DO NOT apply electrodes across the head, or across the heart, or on the front of the neck.
- DO NOT stimulate while driving or operating machinery.

NOTE: The life of the electrode varies depending on skin conditions, skin preparation, type of stimulation, storage and climate. Please read the Instructions For Use for the ManaFlexx 2 thoroughly before use.

ManaFlexx 2
Electrode Sheet

Part No: **MFELP10**

Contents:
Electrode Sheet (10 Pads)

Suggested
HCPC: A4595

LOT 123456

REV A

Part No: **MFELP4**

Single Patient Multiple Use

LATEX

℞ only

MD

Consult instructions for use

NOT STERILE

KEEP DRY

UDI

(01)008101136PROOF
(10)123456

Part No: **MFELP10**

Sugg. HCPC: A4595
ManaFlexx 2 Electrode Sheet (10 Pads)

TD-MFELP10-PKG, Rev. 01